

## **Caesium-135**

An aliquot of the sample solution is diluted using ultrapure nitric acid solution and analysed by inductively coupled plasma mass spectrometry (ICP-MS). The instrument is calibrated externally using stable caesium and is run in peak jumping mode. Rhodium is used as the internal standard. When calculating the results an assumption is made that the stable barium isotopes occur in the natural ratio.