

## **Radon-222**

The method used is a toluene extraction of radon from the sample, followed by liquid scintillation counting of the extract. Counting efficiency is determined by counting radon extracted from a traceable radium-226 standard solution.